Sample Report

*This is an actual report. The clients' home image and location have been removed. The readings do not necessarily reflect typical radon values.



Report Prepared By:

Efficiency Detectives LLC 480 W. Burns Valley RD Winona, Minnesota 55987 507-450-5317

License Number:

107718RT

Model Number: 1030 Serial Number: 94790019 Calibration Date: 10/9/2014

Calibration Factors: [1: 3.11] [2: 2.81] [3: 2.82] [4: 2.93] [5: 2.8] [6: 2.76]

Test Summary: Start Time: 10/17/2014 2:10 AM


Units: pCi/l

Hours Delayed: 12 hour(s) Test Duration: 60 hour(s) Measurement Interval: 1 hour(s)

Measurements: 112

Mitigation System is not installed on property.

Overall Average: 18 pCi/l EPA Average: 16.4 pCi/l



Date/Time		pCi/l	Town (P)	Drogg (in Hg)	Humidity(%)	Plage
10/17/2014	2:40 A	_	72.10		47.0	riags 0
10/17/2014	3:10 A		72.10		47.0	0
10/17/2014	3:40 A		72.00		47.0	0
10/17/2014	4:10 A		72.00		47.0	0
10/17/2014	4:40 A		71.80		47.0	0
10/17/2014	5:10 A		71.80		47.0	0
10/17/2014	5:40 A		71.80		47.0	0
10/17/2014	6:10 A		71.60		47.0	0
10/17/2014	6:40 A		71.60		47.0	0
10/17/2014	7:10 A		71.60		47.0	0
10/17/2014	7:40 A		71.40		47.0	0
10/17/2014	8:10 A		71.40		47.0	0
10/17/2014	8:40 A		71.40		47.0	0
10/17/2014	9:10 A		71.40		47.0	0
10/17/2014	9:40 A		71.40		47.0	0
10/17/2014	10:10 A		71.40		47.0	0
10/17/2014	10:40 A		71.20		47.0	0
10/17/2014	11:10 A		71.20		47.0	0
10/17/2014	11:40 A		71.20		47.0	0
10/17/2014	12:10 P		71.20		47.0	0
10/17/2014	12:40 P		71.10		47.0	0
10/17/2014	1:10 P		71.10	29.2	47.0	0
10/17/2014	1:40 P		71.10	29.2	47.0	0
10/17/2014	2:10 P	M 22.1	70.90	29.2	47.0	0
10/17/2014	2:40 P		70.90	29.2	47.0	0
10/17/2014	3:10 P		70.90	29.2	47.0	0
10/17/2014	3:40 P	M 22.4	70.90	29.2	47.0	0
10/17/2014	4:10 P	M 20.7	70.70	29.2	47.0	0
10/17/2014	4:40 P	M 20.0	70.70	29.2	47.0	0
10/17/2014	5:10 P	M 17.9	70.70	29.2	47.0	0
10/17/2014	5:40 P	M 19.1	70.70	29.3	47.0	0
10/17/2014	6:10 P	M 17.2	70.50	29.3	47.0	0
10/17/2014	6:40 P		70.50	29.3	47.0	0
10/17/2014	7:10 P	M 17.8	70.50		47.0	0
10/17/2014	7:40 P	M 15.4	71.10	29.3	47.0	0
10/17/2014	8:10 P	M 16.4	71.60	29.3	47.0	0
10/17/2014	8:40 P	M 18.0	72.00	29.4	47.0	0
10/17/2014	9:10 P	M 18.1	72.10	29.4	48.0	0
10/17/2014	9:40 P		72.10		47.0	0
10/17/2014	10:10 P	M 16.7	72.10	29.4	48.0	0
10/17/2014	10:40 P	M 18.2	72.10	29.4	48.0	0
10/17/2014	11:10 P	M 17.5	72.00	29.4	47.0	0
10/17/2014	11:40 P	M 17.9	72.00	29.4	47.0	0
10/18/2014	12:10 A		71.80		47.0	0
10/18/2014	12:40 A	M 15.5	71.80	29.4	47.0	0
10/18/2014	1:10 A	M 14.9	71.60	29.4	47.0	0
10/18/2014	1:40 A	M 16.0	71.60	29.4	47.0	0
10/18/2014	2:10 A	M 15.6	71.40	29.4	47.0	0
10/18/2014	2:40 A	M 16.5	71.40	29.4	47.0	0
10/18/2014	3:10 A	M 17.2	71.20	29.4	47.0	0
10/18/2014	3:40 A	M 15.6			47.0	0
10/18/2014	4:10 A	M 17.2	71.10	29.5	47.0	0

Date/Time			pCi/l	Temp(F) I	Press(inHor)	Humidity(%)	Flags
10/18/2014	4:40	AM	15.9	71.10	29.5	47.0	0
10/18/2014	5:10		18.3	71.10	29.5	47.0	ō
10/18/2014	5:40		14.6	70.90	29.5	47.0	ō
10/18/2014	6:10		15.8	70.90	29.5	47.0	ō
10/18/2014	6:40		14.4	70.70	29.5	47.0	ō
10/18/2014	7:10		15.1	70.70	29.5	47.0	ō
10/18/2014	7:40		15.3	70.50	29.5	47.0	ō
10/18/2014	8:10		13.9	70.50	29.5	47.0	ō
10/18/2014	8:40		14.0	70.50	29.6	47.0	ō
10/18/2014	9:10		15.6	70.30	29.6	47.0	0
10/18/2014	9:40		10.9	70.30	29.6	46.0	o
10/18/2014	10:10		12.3	70.20	29.6	46.0	0
10/18/2014	10:40		12.5	70.20	29.6	46.0	0
10/18/2014	11:10		10.9	70.20	29.6	46.0	0
10/18/2014	11:40		9.40	70.00	29.6	46.0	0
10/18/2014	12:10		10.1	70.00	29.6	46.0	0
10/18/2014	12:40		10.3	70.00	29.6	46.0	0
10/18/2014	1:10		7.40	70.30	29.6	46.0	ō
10/18/2014	1:40	PM	8.80	70.70	29.6	46.0	0
10/18/2014	2:10	PM	9.60	70.90	29.6	46.0	0
10/18/2014	2:40	PM	9.50	71.20	29.6	46.0	0
10/18/2014	3:10	PM	10.0	71.60	29.6	46.0	0
10/18/2014	3:40	PM	10.1	72.00	29.6	46.0	0
10/18/2014	4:10	PM	11.1	72.10	29.6	46.0	0
10/18/2014	4:40	PM	11.7	72.30	29.6	46.0	0
10/18/2014	5:10	PM	12.5	72.50	29.6	46.0	0
10/18/2014	5:40	PM	13.2	72.50	29.6	46.0	0
10/18/2014	6:10	PM	11.5	72.50	29.6	46.0	0
10/18/2014	6:40	PM	14.3	72.50	29.6	46.0	0
10/18/2014	7:10	PM	14.3	72.30	29.6	46.0	0
10/18/2014	7:40	PM	15.0	72.10	29.6	46.0	0
10/18/2014	8:10	PM	15.1	72.10	29.6	46.0	0
10/18/2014	8:40	PM	15.0	72.00	29.6	45.0	0
10/18/2014	9:10	PM	15.8	71.80	29.6	45.0	0
10/18/2014	9:40	PM	13.4	71.80	29.6	45.0	0
10/18/2014	10:10	PM	14.2	71.60	29.6	45.0	0
10/18/2014	10:40	PM	13.7	71.60	29.6	45.0	0
10/18/2014	11:10	PM	12.0	71.40	29.6	45.0	0
10/18/2014	11:40	PM	12.9	71.40	29.6	44.0	0
10/19/2014	12:10	AΜ	11.2	71.20	29.6	44.0	0
10/19/2014	12:40	AΜ	13.9	71.20	29.6	44.0	0
10/19/2014	1:10	AΜ	13.4	71.10	29.6	44.0	0
10/19/2014	1:40	ΑM	14.0	71.10	29.6	44.0	0
10/19/2014	2:10		12.1	70.90	29.6	44.0	0
10/19/2014	2:40	AΜ	12.8	70.90	29.5	44.0	0
10/19/2014	3:10	ΑM	12.9	70.70	29.6	44.0	0
10/19/2014	3:40		12.6	70.50	29.5	43.0	0
10/19/2014	4:10		12.2	70.50	29.5	43.0	0
10/19/2014	4:40		9.90	70.30	29.5	43.0	0
10/19/2014	5:10		10.5	70.30	29.5	43.0	0
10/19/2014	5:40		11.8	70.20	29.5	43.0	0
10/19/2014	6:10		9.90	70.20	29.5	43.0	0
10/19/2014	6:40		10.5	70.00	29.5	43.0	0
10/19/2014	7:10	AM	11.2	70.00	29.5	43.0	0

Over All Average:18 pCi/l EPA Average:16.4 pCi/l

Date/Time			pCi/l	Temp(F)	Press(inHg)	Humidity(%)	Flags
10/19/2014	7:40	AM	10.7	69.80	29.5	43.0	0
10/19/2014	8:10	AM	8.50	69.80	29.5	43.0	0
10/19/2014	8:40	AM	7.80	69.80	29.5	43.0	0
10/19/2014	9:10	AM	7.70	69.80	29.5	43.0	0
10/19/2014	9:40	AM	9.80	69.80	29.4	43.0	0
10/19/2014	10-10	ΔM	9.50	69 60	29.4	43 0	0

Over All Average:18 pCi/l EPA Average:16.4 pCi/l

Radon Risk Information

Radon is the second leading cause of lung cancer after smoking. The US EPA and Surgeon General strongly recommend taking further action when a homes radon test results are 4.0 pCi/l or greater. The concentration of radon in the home is measured in picocuries per liter of air (pCi/l). Radon levels less than 4.0 pCi/l still pose some risk and in many cases may be reduced. If the radon level in the home is between 2.0 and 4.0 pCi/l, the EPA still recommends that you consider fixing the home. The average indoor radon level is estimated to be about 1.3 pCi/l; roughly 0.4 pCi/l of radon is normally found in the outside air. The higher the home radon level, the greater the health risk. Even homes with very high radon levels can be reduced to below 4.0 pCi/l and many homes can be reduced to 2.0 pCi/l or less.

Understanding Time-Sensitive Testing Protocols

It is necessary to fix the home when a single test averages 4.0 pCi/l or more. It is a good idea to fix the home when a single test averages between 2.0 and 4.0 pCi/l. If a test result averages less than 4.0 pCi/l, it is recommended to confirm the low result by testing again at least every two years and whenever significant changes to the home structure or mechanical systems occur. Test during different seasons and different weather conditions to reduce your risk of exposure.